Alaris® GS Syringe Pump

Directions For Use

en



CE
0086

CareFusion

Contents

Page

Introduction 2
About This Manual 2
Quick Start Guide 2
Features of the Alaris® GS Syringe Pump 3
Controls and Indicators 4
Symbol Definitions 5
Main Display Features 6
Operating Precautions 7
Getting Started 9
Basic Features 14
Alarms and Warnings 16
Configured Options 17
Alaris® GS Syringe Pump Configured Options Record 19
Specifications 20
Compatible Syringes 21
Associated Products 21
Compatible Extension Sets 22
Maintenance 24
Occlusion Pressure Limits 26
IrDA, RS232 and Nurse Call Specification 27
Trumpet Curves and Start-up Curves 28
Products and Spare Parts 29
Service Contacts 30

Introduction

The Alaris® GS Syringe Pump (herein after referred to as "pump") is a fully featured syringe pump suitable for critical care and general infusion applications.

The Alaris® GS Syringe Pump is compatible with a wide range of standard, single-use, disposable Luer-lock syringes. It accepts syringe sizes from 5 ml to 50 ml. See the 'Compatible Syringes' section for a full list of compatible syringes.

Intended Purpose

The Alaris® GS Syringe Pump is intended for use by medical staff for purposes of controlling infusion rate and volume.

Conditions of Use

The Alaris® GS Syringe Pump should only be operated by a clinician competent in use of automated syringe pumps and post-placement management of intravenous catheters.

> ***CareFusion cannot guarantee the continued system accuracy with other manufacturer's syringes as identified in the 'Compatible Syringes' table. Manufacturers may change syringe specification significant to system accuracy without prior notification.***

Indications

The Alaris® GS Syringe Pump is indicated for infusion of therapeutics including:

- analgesics
- antimicrobials
- blood products
- chemotherapy
- subcutaneous
- nutrition

Contraindications

The Alaris® GS Syringe Pumps is contraindicated for:

- enteral therapies
- epidural

About This Manual

The user must be thoroughly familiar with the Alaris® GS Syringe Pump described in this manual prior to use.

All illustrations used in this manual show typical settings and values which may be used in setting up the functions of the pump. These settings and values are for illustrative use only. Where stated, a minimum infusion rate refers to a nominal rate of 1.0ml/h, and an intermediate infusion rate refers to a nominal rate of 5.0ml/h. The complete range of infusion rates, settings and values are shown in the Specifications section.

Quick Start Guide

1. Press the ⏻ button to turn the pump on.
2. **CLEAR SETUP?** - **NO** retains previous data. **YES** clears previous data.
3. Load syringe.
4. Confirm correct size and brand of syringe.
5. Ensure extension set is attached to syringe, but disconnected from patient.

If the PURGE SYRINGE option has been enabled then the prompt to purge screen is displayed and the set can be purged as required.

6. INFUSION RATE - Change rate if necessary using the ⏫⏬ keys.
7. PURGE - Press the ⏩ button followed by the **PURGE** softkey.
8. Connect extension set to the patient access device.
9. Press the ▶ button to start the infusion.

Features of the Alaris® GS Syringe Pump




Rating Plate (see Symbol Definitions for an explanation of the symbols used)
Release lever for Rotating Cam
Medical Device Interface (MDI)
Rotating Cam to lock on to horizontal rectangular bars
IR Communications port
Carrying Handle
Potential Equalisation (PE) connector
Folded Pole Clamp
RS232 Connector (optional)
Extension set hook

Controls and Indicators

Controls:

Symbol	Description
	ON/OFF button - Press once to switch the pump ON. Press and hold down for 3 seconds to switch the pump OFF.
	RUN button - Press to start the infusion. The green LED will flash during infusion.
	HOLD button - Press to put the infusion on hold. The amber LED will be lit while on hold.
	MUTE button - Press to silence alarm for 2 minutes (configurable). Press and hold until 3 beeps are heard for 15 minutes silence.
	PURGE/BOLUS button - Press to access **PURGE** or **BOLUS** soft keys. Press and hold down soft key to operate. **PURGE** the extension set during set up. • Pump is on hold • Extension set is not connected to the patient • Volume Infused (VI) is not added **BOLUS -** fluid or drug delivered at an accelerated rate. • Pump is infusing • Extension set is connected to the patient • VI is added
	OPTION button - Press to access optional features (see Basic Features).
	CHEVRON keys - Double or single for faster/slower increase or decrease of values shown on display.
	**BLANK SOFTKEYS** - Use in conjunction with the prompts shown on the display.

Indicators:

Symbol	Description
	**BATTERY** indicator - When illuminated the pump is running on the internal battery. When flashing the battery power is low with less than 30 minutes of use remaining.
	**AC POWER** indicator - When illuminated the pump is connected to an AC power supply and the battery is being charged.

Symbol Definitions

Labelling Symbols:

Symbol	Description
	Attention (Consult accompanying documents)
	Potential Equalisation (PE) Connector
	RS232/Nurse call Connector (Optional)
	Defibrillation-proof type CF applied part (Degree of protection against electrical shock)
IPX1	Protected against vertically falling drops of water
	Alternating Current
CE 0086	Device complies with the requirements of Council Directive 93/42/EEC as amended by 2007/47/EC.
	Date of Manufacture
	Manufacturer
	Not for Municipal Waste
	Important information
	Fuse Rating
EC REP	Authorised representative in the European Community



Screen Icons

Symbol	Description
	BATTERY icon - Indicates battery charge level to highlight when the battery will require recharging. This feature can be accessed by pressing the ⓘ button, scroll through the options using the ⏫⏶⏷⏬ keys, select **BATTERY ICON** and this icon will be shown.

Operating Precautions

Disposable Syringes and Extension Sets

- Always clamp or otherwise isolate the patient line before unclamping or removing a syringe from the pump. Failure to do so may result in unintended administration.

- This Alaris® GS Syringe Pump has been calibrated for use with single-use disposable syringes. To ensure correct and accurate operation, only use 3 piece Luer-Lock versions of the syringe make specified on the pump or described in this manual. Use of non-specified syringes or extension sets may impair the operation of the pump and the accuracy of the infusion.



- Uncontrolled flow or syphoning may result if the syringe is located incorrectly in the pump, or if it is removed from the pump before the extension set is properly isolated from the patient. Isolation may include closing a tap in the patient line or activating a flow stop clamp.



- Secure the extension set to the pump using the extension set hook at the rear of the pump. This provides protection against accidental dislodging of the syringe from the pump.



- When combining several apparatus and/or instruments with extension sets and other tubing, for example via a 3-way tap, the performance of the pump may be impacted and should be monitored closely.

Mounting the Pump



- The pump must be mounted within 1.0m above or below the patient's heart. The most accurate pressure monitoring in the extension set is achieved when the pump is positioned close to the patients heart level.

- Do not mount the pump in a vertical position with the syringe pointing upwards as this could lead to an infusion of air which may be in the syringe. To protect against the introduction of air the user should regularly monitor the progress of the infusion, syringe, extension line and patient connections and follow the priming procedure specified herein.

Operating Environment

- When using any infusion pump in conjunction with other pumps or devices requiring vascular access, extra care is advised. Adverse delivery of medication or fluids can be caused by the substantial variation in pressures created within the local vascular system by such pumps. Typical examples of those pumps are used during dialysis, bypass or cardiac assist applications.

- This pump is suitable for use in Hospital and clinical environments other than domestic establishments and those directly connected to the public single phase AC mains power supply network that supplies buildings used for domestic purposes. However, it may be used in domestic establishments under the supervision of Medical professionals with additional necessary appropriate measures. (Consult Technical Service Manual, appropriately trained technical personnel or CareFusion for further information).

- This pump is not intended to be used in the presence of a flammable anaesthetic mixture with air or oxygen or nitrous oxide.

Operating Pressure

- This is a positive pressure pump designed to achieve very accurate fluid administration by automatically compensating for resistance encountered in the infusion system.

- The pumping pressure alarm system is not designed to provide protection against, or detection of, IV complications which can occur.

Alarm Conditions

- Several alarm conditions detected by this pump will stop the infusion and generate visual and audible alarms. Users must perform regular checks to ensure that the infusion is progressing correctly and no alarms are operating.

Electromagnetic Compatibility and Interference

- This pump is protected against the effects of external interference, including high energy radio frequency emissions, magnetic fields and electrostatic discharge (for example, as generated by electrosurgical and cauterising equipment, large motors, portable radios, cellular telephones etc.) and is designed to remain safe when unreasonable levels of interference are encountered.
- Therapeutic Radiation Equipment: Do not use the pump in the vicinity of any Therapeutic Radiation Equipment. Levels of radiation generated by the radiation therapy equipment such as Linear Accelerator, may severely affect functioning of the pump. Please consult manufacturer's recommendations for safe distance and other precautionary requirements. For further information, please contact your local CareFusion representative.
- Magnetic Resonance Imaging (MRI): The pump contains ferromagnetic materials which are susceptible to interference with magnetic field generated by the MRI devices. Therefore, the pump is not considered an MRI compatible pump as such. If use of the pump within an MRI environment is unavoidable, then CareFusion highly recommends securing the pump at a safe distance from the magnetic field outside the identified 'Controlled Access Area' in order to evade any magnetic interference to the pump; or MRI image distortion. This safe distance should be established in accordance with the manufacturers' recommendations regarding electromagnetic interference (EMI). For further information, please refer to the product technical service manual (TSM). Alternatively, contact your local CareFusion representative for further guidance.
- Accessories: Do not use any non-recommended accessory with the pump. The pump is tested and compliant with the relevant EMC claims only with the recommended accessories. Use of any accessory, transducer or cable other than those specified by CareFusion may result in increased emissions or decreased pump immunity.
- This pump is a CISPR 11 Group 1 Class A device and uses RF energy only for its internal function in the normal product offering. Therefore, its RF emissions are very low and are not likely to cause any interference with the nearby electronic equipment. However, this pump emits a certain level of electromagnetic radiation which is within the levels specified by IEC/EN60601-1-2 and IEC/EN60601-2-24. If the pump interacts with other equipment, measures should be taken to minimise the effects, for instance by repositioning or relocation.

- In some circumstances the pump may be affected by an electrostatic discharge through air at levels close to or above 15kv; or by radio frequency radiation close to or above 10v/m. If the pump is affected by this external interference the pump will remain in a safe mode; the pump will duly stop the infusion and alert the user by generating a combination of visual and audible alarms. Should any encountered alarm condition persist even after user intervention, it is recommended to replace that particular pump and quarantine the pump for the attention of appropriately trained technical personnel. (Consult Technical Service Manual for further information).

Hazards

- An explosion hazard exists if the pump is used in the presence of flammable anaesthetics. Exercise care to locate the pump away from any such hazardous sources.

- Dangerous Voltage: An electrical shock hazard exists if the pump's casing is opened or removed. Refer all servicing to qualified service personnel.
- When connected to an external power source, a three-wire (Live, Neutral, Earth) supply must be used. If the integrity of the external protective conductor in the installation or its arrangement is in doubt, the pump should be operated from the battery.

- Do not open the RS232/Nurse Call protective covering when not in use. Electrostatic discharge (ESD) precautions are required when connecting RS232/Nurse Call. Touching the pins of the connectors may result in ESD protection failure. It is recommended that all actions must be taken by appropriately trained personnel.



- If this pump is dropped, subjected to excessive moisture, fluid spillage, humidity or high temperature, or otherwise suspected to have been damaged, remove it from service for inspection by a qualified service engineer. When transporting or storing the pump, use original packaging where possible, and adhere to temperature, humidity and pressure ranges stated in the Specifications section and on the outer packaging.

Getting Started

Initial Set-up

Before operating the pump read this Directions For Use manual carefully.

1. Check that the pump is complete, undamaged and that the voltage rating specified on the label is compatible with your AC power supply.
2. Items supplied are:
   - ***Alaris® GS Syringe Pump***
   - ***User Support CD (Directions For Use)***
   - ***AC Power Cable (as requested)***
   - ***Protective Packaging***
3. Connect the pump to the AC power supply for at least 2½ hours to ensure that the internal battery is charged (verify that the is lit).

Language Selection

1. On initial start-up the pump will display the Select Language screen.
2. Select the required language from the list displayed using the keys.
3. Press the **OK** softkey to confirm your selection.

The pump will automatically operate from its internal battery if the pump is switched on without being connected to the AC power supply.

Should the pump fail to perform correctly, replace in its original protective packaging, where possible and contact a qualified service engineer for investigation.

Do not mount the pump with the AC power inlet or the syringe pointing upwards. This could affect the electrical safety in the event of a fluid spill or lead to the infusion of air which may be in the syringe.

Pole Clamp Installation

The pole clamp is fitted to the rear of the pump and will provide secure fixing to vertical I.V. poles of a diameter between 15 and 40 mm.



1. Pull the folded pole clamp towards you and unscrew the clamp to leave enough room for the size of the pole.
2. Place pump around pole and tighten screw until the clamp is secured to the pole.

Ensure the pole clamp is folded away and stored within the recessed area at the rear of the pump before connecting to a Docking Station/ Workstation* or when not in use.

Never mount the pump such that the IV infusion stand becomes top heavy or unstable.

Docking Station/Workstation or Equipment Rail Installation*



The rotating cam can be fitted to the rectangular bar on the Docking Station/Workstation* or the equipment rail measuring 10 by 25 mm.

1. Align the rotating cam on the rear of the pump with the rectangular bar on the Docking Station/Workstation* or the equipment rail.
2. Hold the pump horizontally, push the pump firmly onto the rectangular bar or equipment rail.
   Ensure that the pump 'clicks' securely into position onto the bar.
3. To release, push the release lever and pull the pump forwards.

*Alaris® DS Docking Station and Alaris® Gateway Workstation.

Loading and Confirming a Syringe

Warning: To securely load and confirm a syringe carefully follow the steps below. An incorrect loading of a syringe may result in misidentification of the syringe type and size. If then confirmed, this may lead to significant inaccuracy of the infusion rate and may also affect pump performance.

Only use a syringe of the type stated on the pump or in this manual. Using an incorrect syringe could adversely affect the accuracy of the infusion rate and may also affect pump performance.

When drawing fluid into the syringe, draw enough to compensate for any 'dead space' volume in the extension set and syringe at the end of infusion as this cannot be fully infused.





Place the pump on a stable horizontal surface or secure as described previously.

Prepare, load and prime the single-use disposable syringe and extension set using standard aseptic techniques.

1. Squeeze the finger grips together on the plunger holder and slide the mechanism to the right.
2. Pull the syringe clamp forward and down.



3. Insert the syringe ensuring that the barrel flange is located in the slots on the syringe flange clamp.



> ***To ensure the syringe is loaded correctly, place the barrel flange in the space between the syringe clamp and the syringe flange clamp. This is correct if the syringe remains in position before the syringe clamp is closed.***

4. Lift the syringe clamp until it locks against the syringe barrel.



5. Squeeze the finger grips on the plunger holder and slide the mechanism to the left until it reaches the plunger end.
6. Release the finger grips. Ensure that the plunger grippers are securing the plunger in place and the finger grip returns to its original position.



7. Ensure that the syringe type and size match those displayed on the pump then press **CONFIRM**. If required, the make of syringe can be changed by pressing the **TYPE** softkey.



Note: If the **PURGE SYRINGE** option has been enabled then the prompt to purge screen is displayed and the extension set can be purged as required, however ensure that the extension set is not connected to the patient during this process.

> ***CareFusion recommends to limit the number of configured syringe types and sizes available for selection on the pump.***
>
> ***Secure the extension set using the extension set hook at the rear of the pump. This provides protection against accidental dislodging of the syringe from the pump.***
>
> ***Ensure that both plunger grippers are fully locked onto the plunger flange and the upper finger grip has returned to its original position.***

Starting the Pump

1. Connect the pump to an AC power supply using the AC power cable.

   Press the ⏻ button.

   - The pump will run a short self-test. Ensure that two beeps are activated during this test.
   - Check the display test pattern and ensure that no rows are missing.
   - Check that the displayed time and date are correct.

   Note: A warning - **REPAIRING LOGS**, may be displayed if event log information was not completely stored at the previous power down. This is for information only, the pump will continue to power up as normal.

2. **CLEAR SETUP? -** Answering **NO** will retain all previous rate and volume settings. **YES** will automatically reset the rate and volume settings to zero.

3. LOAD SYRINGE - Load the syringe according to the procedure in this manual.

4. CONFIRM SYRINGE - Check that the syringe type and size being used matches the display. If required, the make of syringe can be changed by pressing the **TYPE** button. Press **CONFIRM** when the correct type and size are shown.

If the PURGE SYRINGE option has been enabled then the prompt to purge screen is displayed and the set can be purged as required.

5. INFUSION RATE - Check the rate shown if old patient data has been retained and change the rate if necessary using the ⏫⏬ keys.

6. PURGE (if required) - Press the ⏭ button and then press and hold the **PURGE** softkey until fluid flows and the purging of the syringe extension line is complete. Release the softkey. The volume used during purging will be displayed.

7. CONNECT TO PATIENT - Connect the extension set to the patient access device.

8. START - Press ▶ to commence operation. The AMBER STOP light will be replaced by the flashing GREEN START light to indicate that the pump is in operation.

9. STOP - Press ⏹ to halt the operation. The amber light will replace the green light.

Basic Features

Purge

The button allows the delivery of a limited volume of fluid in order to purge the extension set prior to being connected to a patient or after changing a syringe.

1. Press the button when the pump is not infusing. Ensure that the extension set is not connected to the patient.
2. Press and hold the **PURGE** softkey until fluid flows and the purging of the IV infusion set is complete. The volume used during purging will be displayed, but it is not added to the volume infused.
3. When purging is complete release the **PURGE** softkey. Press the **QUIT** softkey to exit back to the main display.

The pump will not purge if the rate lock has been enabled. During PURGE the pressure limit alarms are temporarily increased to their maximum level.

Bolus Infusion

The bolus feature can be used at the start of an infusion or during an infusion.

If the bolus volume reaches the set limit the bolus will stop and the pump will revert to infuse at the set rate.

1. During infusion press the button once to display the bolus screen.
2. Use the keys to set the bolus rate required.
3. To deliver the bolus press the **BOLUS** softkey. During the bolus the volume being infused is displayed.
4. When the desired bolus has been delivered, release the softkey. The bolus volume is added to the total volume infused.

A bolus cannot be administered if the rate lock is enabled or if the feature is disabled in General Options. During BOLUS the pressure limit alarms are temporarily increased to their maximum level.

Clear Volume

This option enables the volume infused to be cleared.

1. Press the **VOLUME** softkey to display the **CLEAR VOLUME** option.
2. Press the **YES** softkey to clear the volume. Press the **NO** softkey to retain the volume.

Selecting YES resets the volume infused in the 24H LOG option.

Rate Lock

If Rate Lock is enabled, when the infusion rate has been set and the infusion started (or following a bolus infusion) the rate lock prompt will appear on the main display.

To select the rate lock function press the **YES** softkey. Press the **NO** softkey if the rate lock is not required.

When rate lock is enabled, the following are unavailable:

- ***Bolus / purge***
- ***Switching the pump off***

To disable the rate lock if selected:

1. Press the button to access the options menu.
2. Select the **UNLOCK RATE** option using the keys and press the **OK** softkey.

To enable the rate lock if not selected:

1. Press the button to access the options menu.
2. Select **RATE LOCK** and press the **OK** softkey.

Basic Features (Continued)

Pressure Display

This feature can be enabled / disabled within the General Options menu (see 'Configured Options'). When enabled, a Pressure Icon appears on the display indicating the current pressure level and the pressure alarm level. The pressure alarm level can be set within the General Options menu.

? 24 Hour Log

This option allows the 24 hour log of volume infused to be reviewed.

1. Press the ? button to access the options menu.
2. Select the **24H LOG** option using the ⏶⏷ keys and press the **OK** softkey.

The display shows the hourly volume infused. The volume infused shown in brackets is the total volume infused since the volume was last cleared. See example below:

07:48 - 08:00 4.34ml (4.34ml)
08:00 - 09:00 2.10ml (6.44ml)
09:00 - 10:00 2.10ml (8.54ml)
VOLUME CLEARED

3. Press the **QUIT** softkey to exit the log.

? Event Log

This option allows the event log to be reviewed. It can be enabled/disabled.

1. Press the ? button to access the options menu.
2. Select the **EVENT LOG** option using the ⏶⏷ keys and press the **OK** softkey.
3. Scroll through the log using the ⏶⏷ keys. Press the **QUIT** softkey to exit the log.

Alarms and Warnings

Alarms are indicated by a combination of an audible alarm, flashing alarm indicator and a descriptive message in the display.

1. First press the ⊗ button to silence the alarm for a maximum of 2 minutes*, then check the display for an alarm message. Press **CANCEL** to cancel the alarm message.
2. If the infusion has stopped, rectify the cause of the alarm then press the ⏻ button to resume the infusion.

If the pump initiates a safety processor alarm condition (an audible high pitched continuous shrill accompanied with a red alarm indicator) and there is no error message displayed on the pump, remove the pump from service for examination by a qualified service engineer.

Display	*Description and Troubleshooting Guide*
DRIVE DISENGAGED	The drive system has been disengaged during operation. Check the finger grips and the position of the syringe.
OCCLUSION	Excessive pressure measured at the syringe plunger exceeding the alarm limit. Identify and remove the cause of the blockage in the drive, syringe, or administration system before restarting the infusion.
CHECK SYRINGE	Incorrect size of syringe has been fitted, the syringe has not been positioned correctly or has been disturbed during operation. Check the syringe location and the position.
BATTERY LOW	Battery charge low with 30 minutes operation remaining. Battery indicator will flash and after 30 minutes a continuous audible alarm will indicate that the battery is exhausted. Reconnect to the AC power supply to continue operation and charge the internal battery.
BATTERY EMPTY	The internal battery is exhausted. Connect the pump to the AC power supply.
NEAR END OF INFUSION	The pump is nearing the end of the infusion. This value can be configured.
END OF INFUSION	The pump has reached the end of the infusion. A pre-set volume will remain in the syringe to minimise the risk of the infusion of air bubbles into the set. This value can be configured.
AC POWER FAIL	AC Power has been disconnected and the pump is operating on battery power, if this occurs when the pump is infusing the message "**INFUSION CONTINUES**" will be displayed. Reconnect AC power supply or press the ⊗ button to silence the alarm and continue with battery operation. The alarm will automatically cancel if the AC power supply is reconnected.
Error Code and Message	The alarm system has detected an internal malfunction. Note the malfunction code. Remove pump from service for examination by a qualified service engineer.
ATTENTION *(with "3 Beeps")*	Three beeps will sound if the pump has been left ON for more than 2 minutes* (referred to as **CALLBACK** in the log) without starting the operation. Press the ⊗ button to silence the alarm for a further 2 minutes*. Alternatively press and hold down the ⊗ button and wait for 3 beeps in succession, this will put the warning alarm on standby for 15 minutes.

Alarm Indicator Colour	*Alarms indicated*
AMBER	AC POWER FAIL; NEAR END OF INFUSION; ATTENTION; BATTERY LOW.
RED	All others.

*Configurable option.

Configured Options

This menu comprises a list of options which are configurable by the user.

1. Turn the pump **OFF**.
2. Whilst holding down the ⏻ button turn the pump **ON**.
3. The main display will show **000**. Enter the access code for Configured Options using the ⏶⏷ keys, pressing **NEXT** to move through the digits. A full list of access codes can be found in the Technical Service Manual.
4. When the complete code shows on screen, press **OK** to enter. The Configured Options menu will be displayed.

General Options

1. Select **GENERAL OPTIONS** from the menu using the ⏶⏷ keys and press the **OK** softkey.
2. Select the option you wish to enable/disable or adjust and press the **MODIFY** softkey.
3. When all the desired modifications have been carried out press the **QUIT** softkey.
4. Either select the next configuration option from the menu or turn the pump **OFF**, returning it to operation as required.

NURSE CALL FITTED	Enables Nurse Call (hardware option).
NURSE CALL INVERT	When enabled, the nurse call output is inverted.
RS232 SELECTED	Sets the pump's communications to use RS232 (hardware option).
NEOI WARNING	Sets the Near End Of Infusion warning time, as time left to End Of Infusion.
EOI POINT	Sets the End Of Infusion point.
KVO AT EOI	When enabled the pump will switch to running at the KVO rate when EOI is reached.
KVO RATE	Sets the Keep Vein Open (KVO) rate at which the pump will operate if KVO at EOI is enabled.
BACK OFF	When enabled the motor will reverse to relieve line pressure when an occlusion occurs.
AUTO SAVE	When disabled the infusion information is cleared on power up.
RATE LOCK	When enabled the rate can be locked to prevent unwanted changes of the set infusion rate.
QUIET MODE	When enabled the button beeps are muted.
AC FAIL	When enabled the AC Power Failure Alarm will sound if the AC power is disconnected.
PRESSURE DISPLAY	Enables / disables the Pressure Icon on the main display.
PRESSURE DEFAULT	Sets the default occlusion alarm level.
CAP RATE	Sets the maximum value for infusion rate.
PURGE RATE	Sets the purge rate.
PURGE VOLUME LIMIT	Sets the maximum permissible purge volume.
PURGE SYRINGE	Prompt to purge syringe after confirmation.
BOLUS	Enables / disables the bolus feature.
DEFAULT BOLUS	Sets the default bolus rate.
CAP BOLUS RATE	Sets the maximum value for bolus rate.
BOLUS VOL LIMIT	Sets the maximum permissible bolus volume.
MANUAL BOLUS	Volume infused will be increased if plunger is manually moved in and syringe remains confirmed.
CALL BACK TIME	Adjusts the time for the pump to sound the call back alarm.
EVENT LOG DISPLAY	Enables / disables the event log.
BATTERY ICON	Enables / disables the Battery Icon feature availability via the ? button.
AUDIO VOLUME	Sets the alarm volume of the pump at high, medium or low.
AUTO NIGHT MODE	Backlight dims between hours 21:00 and 06:00.

Configured Options (Continued)

Clock Set

1. Select **CLOCK SET** from the Configured Options menu using the ⏫🔼🔽⏬ keys and press the **OK** softkey.
2. Use the ⏫🔼🔽⏬ keys to adjust the date displayed, pressing the **NEXT** softkey to access the next field.
3. When the correct time and date are displayed press the **OK** softkey to return to the Configured Options menu.

Hospital Name

This option allows the user to programme in the name of the hospital, ward or department. This will appear during the power-up display sequence.

1. Select **HOSPITAL NAME** from the Configured Options menu using the ⏫🔼🔽⏬ keys and press the **OK** softkey.
2. Use the keys to adjust the character displayed, pressing **NEXT** to access the next position.
3. When the correct name is displayed press **OK** to return to the Configured Options menu.

Enable Syringes

This option is used to pre-configure the type and size of syringe permitted for use on the pump. Select all possible syringes which may be used and disable any that should not be used.

1. Select **ENABLE SYRINGES** from the Configured Options menu using the ⏫🔼🔽⏬ keys and press the **OK** softkey.
2. Use the ⏫🔼🔽⏬ keys to scroll through the list of syringes, pressing **MODIFY** to enable/disable a syringe brand and individual models within the brand.
3. When all modifications are complete press **QUIT** to return to the Configured Options menu.

Language

This option is used to set the language of messages shown on the pump display.

1. Select **LANGUAGE** from the Configured Options menu using the ⏫🔼🔽⏬ keys and press the **OK** softkey.
2. Use the ⏫🔼🔽⏬ keys to select the language.
3. When the desired language has been selected press **SELECT** softkey to return to the Configured Options menu.

Contrast

This option is used to set the contrast on the pump display.

1. Select **CONTRAST** from the Configured Options menu using the ⏫🔼🔽⏬ keys and press the **OK** softkey.
2. Use the ⏫🔼🔽⏬ keys to select a contrast ratio value. The contrast of the display will change when scrolling through the numbers.
3. When the desired value has been reached press the **OK** softkey to return to the Configured Options menu.

Alaris® GS Syringe Pump Configured Options Record

General Options

Enter the pump-specific information for your records on a copy of this page.

Option	Default		Range	Setting
Software Version	**1.5.10 and 2.0.0**	**1.9.x and 2.3.x and above**		
NURSE CALL FITTED	Disabled	Disabled	Enabled/Disabled	
NURSE CALL INVERT	Disabled	Disabled	Enabled/Disabled	
RS232 SELECTED	Disabled	Disabled	Enabled/Disabled	
NEOI WARNING	1min	5mins	1min - 15mins	
EOI POINT	1.0%	1.0%	0.1% - 5% of syringe volume	
KVO AT EOI	Enabled	Enabled	Enabled/Disabled	
KVO RATE	1.0ml/h	1.0ml/h	0.1ml/h - 2.5ml/h	
BACK OFF	Disabled	Enabled	Enabled/Disabled	
AUTO SAVE	Enabled	Enabled	Enabled/Disabled	
RATE LOCK	Disabled	Disabled	Enabled/Disabled	
QUIET MODE	Disabled	Disabled	Enabled/Disabled	
AC FAIL	Enabled	Enabled	Enabled/Disabled	
PRESSURE DISPLAY	Disabled	Enabled	Enabled/Disabled	
PRESSURE DEFAULT	L3	L3	L0 - L10 (50mmHg - 1000mmHg)	
CAP RATE	Max infusion rate	200ml/h	1.0ml/h - 200ml/h	
PURGE RATE	200ml/h	200ml/h	100ml/h - 500ml/h	
PURGE VOLUME LIMIT	2.0ml	2.0ml	0.5ml - 5.0ml	
PURGE SYRINGE		Disabled	Enabled/Disabled	
BOLUS	Enabled	Enabled	Enabled/Disabled	
DEFAULT BOLUS	Max bolus rate	500ml/h	10ml/h - 500ml/h	
CAP BOLUS RATE	Max bolus rate	500ml/h	10ml/h - 500ml/h	
BOLUS VOL LIMIT	5.0ml	5.0ml	0.5ml (0.1ml)* - 25.0ml	
MANUAL BOLUS		Disabled	Enabled/Disabled	
CALL BACK TIME		2mins	0.1mins - 15mins	
EVENT LOG DISPLAY	Disabled	Enabled	Enabled/Disabled	
BATTERY ICON		Enabled	Enabled/Disabled	
AUDIO VOLUME	Medium	Medium	Low, medium, high	
AUTO NIGHT MODE	Enabled	Enabled	Enabled/Disabled	

* For software versions 1.9.x and 2.3.x and above

Syringes Enabled

Make	Size(s)	Make	Size(s)

Hospital Name	Serial No.	Software Version
Approved by	Configured by	
Date	Date	

Specifications

Infusion Specifications -

Maximum infusion rate can be set as part of the configuration.

0.1ml/h - 150ml/h	5ml syringes
0.1ml/h - 200ml/h	All other syringes

The Volume Infused range is 0.0ml - 9990ml.

Bolus Specifications -

Maximum Bolus rates can be set as part of the configuration. Bolus rates are user adjustable, in increments of 10ml/h.

10 ml/h - 150ml/h	5ml syringes
10 ml/h - 300ml/h	10ml syringes
10 ml/h - 500ml/h	All other syringes

The bolus volume limit can be set as part of the configuration.

Minimum: 0.5ml (0.1ml - v2.3.x and above or v1.9.x);
Maximum 25.0ml
Increments of 0.1ml; default 5.0ml

During BOLUS the pressure limit alarms are temporarily increased to their maximum level.

Critical Volume -

The bolus which can occur in the event of a single internal fault condition with a 50 ml syringe is :

Maximum Overinfusion - 0.5ml

Purge Specifications -

The purge rate is limited to the maximum rate for the syringe and can be set as part of the configuration.

100ml/h - 500ml/h.

The purge volume range is 0.5ml - 5ml.

During PURGE the pressure limit alarms are temporarily increased to their maximum level.

Keep Vein Open (KVO) Rate -

0.1 ml/h - 2.5ml/h.

End Of Syringe Rate -

Stop, KVO (0.1ml/h to 2.5ml/h), or set rate if lower than KVO.

Near End Of Infusion Alarm -

1min - 15min to end of infusion, or 10% of syringe volume, whichever is smaller.

End Of Infusion (EOI) Alarm -

0.1% - 5% of syringe volume

Electrical Classification -

Class I product. Continuous Mode Operation, Transportable

Maximum Pumping Pressure Limit -

Highest alarm level 1000mmHg (nominal at L-10)

Occlusion Accuracy (% of full scale)* -

	Pressure mmHg			
	L-0 approx. 50 mmHg	L-3 approx. 300 mmHg	L-5 approx. 500 mmHg	L-10 approx. 1000 mmHg
Temp. 23°C	±18%	±21%	±23%	±28%

* - Using most common 50ml syringes under normal conditions (95% confidence / 95% of pumps).

System Accuracy -

Volumetric Mean +/- 2% (nominal).

Derating -

Temperature +/- 0.5% (5 - 40°C)

High Rates +/-2.0% (rates > syringe volume/h eg. >50ml/h in a 50ml syringe.)

Important: System accuracy is +/-2% typical by volume as measured using the trumpet curve test method defined in IEC/EN60601-2-24 at rates of 1.0ml/h (23°C) and above when the pump is used with the recommended syringes. Caution: Infusion volume accuracy may be compromised at rates below 1.0ml/h. Differences in factors such as size and plunger force in compatible syringes can cause variations in accuracy and trumpet curves. See also trumpet curves section in this manual.

Battery Specifications -

Rechargeable sealed NiMH. Automatically charges when the pump is connected to AC power.

Mean Time To Battery Empty from fully charged @ 5ml/h and 23°C ± 2°C under normal conditions is 6 hours*

*95% lower confidence interval of 5 hours 50 minutes

Charging takes 2½ hours from discharge to 90% charge.

Memory Retention -

The electronic memory of the pump will be retained for more than 6 months when not powered up.

Fuse Type -

2 x T 1.25A, slow blowing.

AC Power Supply -

115 - 230VAC, 50 - 60Hz, 20VA (nominal).

Dimensions -

310 mm (w) x 121 mm (h) x 200 mm (d). Weight: 2.7 kg (excluding power cable).

Protection against fluid ingress -

IPX1 - Protected against vertically falling drops of water.

Alarm Conditions -

Drive Disengaged	Occlusion
Check Syringe	Battery Low / Battery Empty
Near End Of Infusion	End of Infusion
AC Power Fail	Internal Malfunction
Attention (Nurse Callback)	

Environmental Specifications -

Operating Temperature	**+5°C - +40°C**
Operating Relative Humidity	**20% - 90%**
Operating Atmospheric Pressure	**700hPa - 1060hPa**
Transport and Storage Temperature	**-30°C - +50°C**
Transport and Storage Relative Humidity	**10% - 95%**
Transport and Storage Atmospheric Pressure	**500hPa - 1060hPa**

Electrical/Mechanical Safety -

Complies with IEC/EN60601-1 and IEC/EN60601-2-24.

EMC -

Complies with IEC/EN60601-1-2 and IEC/EN60601-2-24.

Compatible Syringes

The pump is calibrated and labelled for use with single-use disposable Luer-lock syringes. Only use the size and type of syringe specified on the pump display. The full list of permitted syringe models is dependent on the software version of the pump.

	5ml	10ml	20ml	30ml	50ml
IVAC®					✓
AstraZeneca					✓
B Braun Omnifix	✓	✓	✓	✓	✓
B Braun Perfusor			✓		✓
BD Perfusor					✓
BD Plastipak	✓	✓	✓	✓	✓
BD Precise			✓		✓
Codan		✓	✓	✓	✓
Codan Perfusion					✓
Fresenius Injectomat		✓			✓
Monoject**	✓	✓	✓	✓	✓
Nipro	✓		✓	✓	✓
Pentaferte	✓	✓	✓		✓
Rapiject*					✓
Terumo	✓	✓	✓	✓	✓

* - The Rapiject 50ml syringe is a specialised syringe with a large diameter barrel. To provide protection against accidental dislodging always ensure the infusion line is secured using the infusion set hook - see Loading a Syringe section.
** - ≡TYCO / Healthcare KENDALL - MONOJECT.

Associated Products

The Alaris® DS Docking Station



The Alaris® Gateway Workstation



Compatible Extension Sets

The pump uses standard, single-use, disposable extension sets and syringes with Luer-lock connectors. The user is responsible for verifying the suitability of a product used, if it is not recommended by CareFusion.

Standard Sets

G40015 Standard PVC Syringe Extension Set (150 cm).
Priming Volume: 2.6ml

G40020B Standard PVC Syringe Extension Set (200 cm).
Priming Volume: 1.5ml

G402EP Extension set, luer lock connectors. Kink resistant DEHP free PVC yellow striped tubing. Bore 1mm. Length 200cm.
Priming volume 1.6ml.

Light Protected Sets

G40215 Amber PE Syringe Extension Set (150 cm).
Priming Volume: 1.2ml

G40320 White PVC Syringe Extension Set (200 cm).
Priming Volume: 3.6ml

Low Sorbing Sets

G40615 Polyethylene Syringe Extension Set (150 cm).
Priming Volume: 1.5ml

G40620 Polyethylene Syringe Extension Set (200 cm).
Priming Volume: 2ml

G40720 Polyethylene Lined Syringe Extension Set with clamp. (200 cm).
Priming Volume: 1.5ml



04105010509 Polyethylene Syringe Extension Set (100 cm).
Priming Volume: 1ml

For availability please contact your local CareFusion representative because new sets are continuously being developed for our customers.

It is recommended that extension sets are changed in accordance with the Directions for Use.
Carefully read the Directions For Use supplied with the extension set prior to use.

Please note these drawings are not to scale

Compatible Extension Sets (Continued)

The pump uses standard, single-use, disposable extension sets and syringes with Luer-lock connectors. The user is responsible for verifying the suitability of a product used, if it is not recommended by CareFusion.

Patient Controlled Analgesia (PCA) Sets

30822 PVC Syringe Extension Set with clamp (152 cm).
Priming Volume: 0.5ml

30832 PVC 'Y' Syringe Extension Set with back check valve and 2 clamps (178 cm).
Priming Volume: 1.5ml



30842E PVC Syringe Extension Set with back check valve, SmartSite® Needle-Free Valve Port and clamp (30 cm).
Priming Volume: 1.4ml



30852 PVC 'Y' Syringe Extension Set with anti-siphon valve, back check valve and 2 clamps (183 cm).
Priming Volume: 1.8ml



30862 PVC Syringe Extension Set with anti-siphon valve and clamp (156 cm).
Priming Volume: 0.6ml



04102215162 PVC Syringe Extension Set with rotating luer. (150 cm).
Priming Volume: 2.9ml



04100010162 PVC Syringe Extension Set (105 cm).
Priming Volume: 7.2ml

For availability please contact your local CareFusion representative because new sets are continuously being developed for our customers.

It is recommended that extension sets are changed in accordance with the Directions for Use.
Carefully read the Directions For Use supplied with the extension set prior to use.

Please note these drawings are not to scale

Maintenance

Routine Maintenance Procedures

To ensure that this pump remains in good operating condition, it is important to keep it clean and carry out the routine maintenance procedures described below.

Interval	*Routine Maintenance Procedure*
As per Hospital Policy	Thoroughly clean external surfaces of the pump before and after prolonged period of storage.
Each usage	1. Inspect AC power supply plug and cable for damage.
	2. Inspect case, keypad and plunger for damage.
	3. Check Start up self test operation is correct.
Before the transfer of the pump to a new patient and as required	Clean the pump by wiping over with a lint-free cloth lightly dampened with warm water and a standard disinfectant / detergent solution.

If the pump is dropped, damaged, subjected to excessive moisture or high temperature, immediately take it out of service for examination by a qualified service engineer.

All preventative and corrective maintenance and all such activities shall be performed at a compliant work place in accordance with the information supplied. CareFusion will not be responsible should any of these actions be performed outside the instructions or information supplied by CareFusion. For Preventative and Corrective Maintenance instructions please refer to the Technical Service Manual (TSM).

All servicing should only be performed by a qualified service engineer with reference to the TSM.

Please refer to Technical Service Manual for calibration procedures. The units of measurement used in the calibration procedure are standard SI (The International System of Units) units.

Battery Operation

The internal rechargeable battery allows continued operation when the AC power is unavailable, for example during patient transfer or AC power failure. Mean Time To Battery Empty from fully charged @ 5ml/h and 20°C under normal conditions is 6 hours*. From the battery low alarm it will take about 2½ hours to 90% charge when reconnected to the AC power supply, whether the pump is in use or not.

The battery is maintenance free, sealed Nickel Metal Hydride and requires no routine servicing. However, to achieve optimum operation, ensure that the battery is fully recharged after full discharge, before storage, and at regular 3 month intervals during storage.

It is recommended that only a qualified service engineer replaces the battery. For further information regarding the replacement of batteries refer to the Technical Service Manual.

The battery pack used in this Alaris® Syringe Pump is manufactured by CareFusion and includes a proprietary PCB (printed circuit board) designed specifically for the Alaris® Syringe Pump, and in conjunction with Alaris® Syringe Pump software, controls battery use, charge and temperature. Any use of battery packs that are not manufactured by CareFusion in the Alaris® Syringe Pump is at your sole risk, and CareFusion does not provide any warranty for or endorsement on any battery packs that are not manufactured by CareFusion. CareFusion's product warranty shall not apply in the event the Alaris® Syringe Pump has suffered damage or premature wear, or malfunctions or otherwise operates incorrectly, as a result of use with a battery pack that is not manufactured by CareFusion.

*95% lower confidence interval of 5 hours 50 minutes

Cleaning and Storage

Before the transfer of the pump to a new patient and periodically during the use, clean the pump by wiping over with a lint-free cloth lightly dampened with warm water and a standard disinfectant / detergent solution.

Recommended cleaners are:

Brand	Concentration
Hibiscrub	20% (v/v)
Virkon	1% (w/v)

Do not use the following disinfectant types:

- Disinfectants which are known to be corrosive to metals must not be used, which include:
  - NaDcc (such as Presept),
  - Hypochlorites (such as Chlorasol),
  - Aldehydes (such as Cidex),
  - Cationic Surfactants (such as Benzalkonium Chloride).
- Use of Iodine (such as Betadine) will cause surface discoloration.
- Concentrated Isopropyl alcohol based cleaners will degrade plastic parts.

The syringe and extension sets are disposable single use items and should be discarded after use according to their manufacturers' instructions.

If the pump is to be stored for an extended period it should be first cleaned and the internal battery fully charged. Store in a clean, dry atmosphere at room temperature and, if available, employ the original packaging for protection.

Once every 3 months during storage, carry out functional tests as described in the Technical Service Manual and ensure that the internal battery is fully charged.

Before cleaning always switch OFF and disconnect from the AC power supply. Never allow liquid to enter the casing and avoid excess fluid build up on the pump. Do not use aggressive cleaning agents as these may damage the exterior surface of the pump. Do not steam autoclave, ethylene oxide sterilise or immerse this pump in any fluid.

Disposal

Information on Disposal for Users of Waste Electrical and Electronic Equipment

This symbol on the product and/or accompanying documents means that used electrical and electronic products should not be mixed with household waste.

If you wish to discard electrical and electronic equipment, please contact your CareFusion affiliate office or distributor for further information.

Disposing of this product correctly will help to save valuable resources and prevent any potential negative effects on human health and the environment which could otherwise arise from inappropriate waste handling.

Information on Disposal in Countries outside the European Union

This symbol is only valid in the European Union. The product should be disposed of taking environmental factors into consideration. To ensure no risk or hazard, remove the internal rechargeable battery and the Nickel Metal Hydride battery from the control board and dispose of as outlined by the local country regulations. All other components can be safely disposed of as per local regulations.

Occlusion Pressure Limits

Time to alarm following occlusion is achieved in less than 30 minutes at rates of 1 ml/h and higher by the appropriate selection of occlusion levels.
The following graphs show the typical values for time to alarm and bolus volume that can be expected in the event of an occlusion when the BD Plastipak 50 ml syringe is selected with a G40020B standard extension set.



Tests at low alarm levels may alarm immediately - the force at these levels is commonly less than the friction in the syringe (with no additional fluid pressure). The result is that the pressure relating to the low forces will be less than the nominal quoted occlusion pressure.

Bolus volume following occlusion will be minimised by the back off feature if enabled. The back off will reduce the line pressure by removing the volume stored in the occluded line and deducting this volume from the volume infused.

IrDA, RS232 and Nurse Call Specification

IrDA / RS232 / Nurse Call Feature

The RS232 / Nurse call feature is an optional feature on Alaris® Syringe Pumps. It allows the pump to be monitored remotely and/or controlled via a suitable central monitoring or computer system.

When the pump is started by a command from the serial interface, communication must take place over the serial interface, a communication must take place every 15 seconds or the pump will alarm, display communications failure and stop infusing. This failure protects against failure of the communications, including the removal of the RS232 cable.

The nurse call interface provides a remote backup to the internal audible alarm. It should not be relied upon to replace monitoring of the internal alarm.

Refer to the Technical Service Manual for further information regarding the RS232 interface. Since it is possible to control the syringe pump using the RS232 interface at some distance from the pump and hence remote from the patient, responsibility for the control of the pump is vested in the software run on the computer control system.

The assessment for the suitability of any software used in the clinical environment to control or receive data from the pump lies with the user of the equipment. This software should include detection of the disconnection or other failure of the RS232 cable. The protocol is detailed in the Technical Service Manual and is for general information only.

Any connected analogue and digital components are required to meet IEC/EN60950 for data processing and IEC/EN60601 for medical devices. Anyone connecting additional devices to the signal input or output is a system configurator and responsible for meeting the requirements of the system standard IEC/EN60601-1-1.

IrDA

Baud Rate	38.4 kBaud
Start Bits	1 Start Bit
Data Bits	8 Data Bits
Parity	No Parity
Stop Bits	1 stop bit

RS232 / Nurse Call Connection Data

Nurse call Specification -

Connector	D Type - 9 Pin
TXD/RXD	EIA RS232-C Standard
TXD Output Voltage Range	Minimum: -5V (mark), +5V (space) Typical: -7V (mark), +7V (space) with 3kΩ load to ground
RXD Input Voltage Range	-30V - +30V max.
RXD Input Thresholds	Low: 0.6V minimum / High: 3.0V maximum
RXD Input Resistance	3kΩ minimum
Enable	Active, Low:-7V to -12V Active, High:+7V to +12V, powers up the isolated RS232 circuitry Inactive: Floating/open circuit, allows isolated RS232 circuitry to power down.
Isolation Socket/Pump	1.5kV (dc, or ac peak)
Baud Rate	38.4 kBaud
Start Bits	1 Start Bit
Data Bits	8 Data Bits
Parity	No Parity
Stop Bits	1 stop bit
Nurse Call Relay Contacts	Pins 1, 8 + 9, 30V dc, 1A rating

Typical Connection Data -

1 Nurse call (Relay) Normally Closed (NC C)

2 Transmit Data (TXD) Output

3 Received Data (RXD) Input

4 Power Input (DSR)

5 Ground (GND)

6 Not used

7 Power Input (CTS)

8 Nurse call (Relay) Normally open (NC O)

9 Nurse call (Relay) Common (NC COM)



Trumpet Curves and Start-up Curves

In this pump, as with all infusion systems, the action of the pumping mechanism and variations in individual syringes cause short-term fluctuations in rate accuracy.

The following curves show typical performance of the system in two ways: 1) the delay in onset of fluid flow when infusion commences (start-up curves), and 2) the accuracy of fluid delivery over various time periods is measured (trumpet curves).

The start-up curves represent continuous flow versus operating time from the start of the infusion. They exhibit the delay in onset of delivery due to mechanical compliance and provide a visual representation of uniformity. Trumpet curves are derived from the second hour of this data. Tests performed per IEC/EN60601-2-24 standard.

Trumpet curves are named for their characteristic shape. They display discrete data averaged over particular time periods or 'observation windows', not continuous data versus operating time. Over long observation windows, short term fluctuations have little effect on accuracy as represented by the flat part of the curve. As the observation window is reduced, short term fluctuations have greater effects as represented by the "mouth" of the trumpet.

Knowledge of system accuracy over various observation windows may be of interest when certain drugs are being administered. Short term fluctuations in rate accuracy may have clinical impact depending on the half-life of the drug being infused, therefore the clinical effect cannot be determined from the trumpet curves alone.

Start-up and trumpet curves may not be indicative of operation under negative pressure.

Differences in factors such as size and plunger force in compatible syringes produced by other manufacturers can cause variations in accuracy and trumpet curves as compared to those represented. Additional curves for compatible syringes are available upon written request.

For applications where flow uniformity is a concern, rates of 1.0 ml/h or above are recommended.

Start-up Trend. BD Plastipak 50ml @ 0.1ml/h



Trumpet Curve. BD Plastipak 50ml @ 0.1ml/h



Start-up Trend. BD Plastipak 50 ml @ 1.0 ml/h



Trumpet Curve. BD Plastipak 50 ml @ 1.0 ml/h



Start-up Trend. BD Plastipak 50 ml @ 5.0 ml/h



Trumpet Curve. BD Plastipak 50 ml @ 5.0 ml/h



Alaris® Infusion System

Range of products in the Alaris® Infusion System product family are:

Part Number	Description
80023UN01	Alaris® GH Syringe Pump
80033UND1	Alaris® CC Syringe Pump
80043UN01	Alaris® TIVA Syringe Pump
80053UN01	Alaris® PK Syringe Pump
80033UND1-G	Alaris® CC Guardrails® Syringe Pump
80023UN01-G	Alaris® GH Guardrails® Syringe Pump
80263UN01-G	Alaris® GP Guardrails® Volumetric Pump
274	Alaris® Transporter
80083UN00-xx[1]	Alaris® DS Docking Station
80203UNS0x-xx[1]	Alaris® Gateway Workstation

[1] For Docking Stations and Workstation contact local customer services representative to obtain configurations availability and part numbers.

Spare Parts

A comprehensive list of spare parts for this pump is included within the Technical Service Manual.

The Technical Service Manual (1000SM00001) is now available in electronic format on the World Wide Web at :-

www.carefusion.co.uk/alaris-technical/

A username and password are required to access our manuals. Please contact local customer services representative to obtain login details.

Part Number	Description
1000SP01122	Internal Battery Pack
1001FAOPT91	AC Power Lead - UK
1001FAOPT92	AC Power Lead - European

Service Contacts

For service contact your local Affiliate Office or Distributor.

AE	***CN***	***GB***	***NZ***
CareFusion, PO Box 5527, Dubai, United Arab Emirates.	CareFusion, 上海代表机构，中国上海市张杨路 500 号， 上海时代广场办事处大楼， A 座，24 层， 邮编：200122。	CareFusion, The Crescent, Jays Close, Basingstoke, Hampshire, RG22 4BS, United Kingdom.	CareFusion, 14B George Bourke Drive, Mt Wellington 1060, PO Box 14-518, Panmure 1741, Auckland, New Zealand
Tel: (971) 4 28 22 842	电话：(86) 21 58368018	Tel: (44) 0800 917 8776	Tel: 09 270 2420 Freephone: 0508 422734
Fax: (971) 4 28 22 914	传真：(86) 21 58368017	Fax: (44) 1256 330860	Fax: 09 270 6285
AU	***DE***	***HU***	***PL***
CareFusion, 3/167 Prospect Highway, PO Box 355 Seven Hills, NSW 2147, Australia.	CareFusion, Pascalstr. 2, 52499 Baesweiler, Deutschland.	CareFusion, Döbrentei tér 1, H-1013 Budapest, Magyarország.	CareFusion, ul. Rzymowskiego 53, 02-697 Warszawa, Polska.
Tel: (61) 1800 833 372	Tel: (49) 931 4972 837	Tel: (36) 1 488 0232 Tel: (36) 1 488 0233	Tel: (48) 225480069
Fax: (61) 1800 833 518	Fax: (49) 931 4972 318	Fax: (36) 1 201 5987	Fax: (48) 225480001
BE	***DK***	***IT***	***SE***
CareFusion, Leuvensesteenweg 248 D, 1800 Vilvoorde, Belgium.	CareFusion, Firskovvej 25 B, 2800 Lyngby, Danmark.	CareFusion, Via Ticino 4, 50019 Sesto Fiorentino, Firenze, Italia.	CareFusion, Hammarbacken 4B, 191 46 Sollentuna, Sverige.
Tel: (32) 2 267 38 99	Tlf. (45)70 20 30 74	Tél: (39) 055 30 33 93 00	Tel: (46) 8 544 43 200
Fax: (32) 2 267 99 21	Fax. (45)70 20 30 98	Fax: (39) 055 34 00 24	Fax: (46) 8 544 43 225
CA	***ES***	***NL***	***US***
CareFusion, 235 Shields Court, Markham, Ontario L3R 8V2, Canada.	CareFusion, Edificio Veganova, Avenida de La Vega, nº1, Bloque 1 - Planta 1, 28108 Alcobendas, Madrid, España.	CareFusion, De Molen 8-10, 3994 DB Houten, Nederland.	CareFusion, 10020 Pacific Mesa Blvd., San Diego, CA 92121, USA.
Tel: (1) 905-752-3333	Tel: (34) 902 555 660	Tel: +31 (0)30 2289 711	Tel: (1) 800 854 7128
Fax: (1) 905-752-3343	Fax: (34) 902 555 661	Fax: +31 (0)30 2289 713	Fax: (1) 858 458 6179
CH	***FR***	***NO***	***ZA***
CareFusion Switzerland 221 Sàrl Critical Care A-One Business Centre Zone d'activitiés Vers-la-Pièce n° 10 1180 Rolle / Switzerland	CareFusion, Parc d'affaire le Val Saint Quentin 2, rue René Caudron 78960 Voisins le Bretonneux France	CareFusion, Solbråveien 10 A, 1383 ASKER, Norge.	CareFusion, Unit 2 Oude Molen Business Park, Oude Molen Road, Ndabeni, Cape Town 7405, South Africa.
Ph.: 0848 244 433	Tél: (33) 1 30 05 34 00	Tel: (47) 66 98 76 00	Tel: (27) (0) 860 597 572 Tel: (27) 21 510 7562
Fax: 0848 244 100	Fax: (33) 1 30 05 34 43	Fax: (47) 66 98 76 01	Fax: (27) 21 5107567

Rev. B

Alaris, Guardrails, SmartSite and IVAC are registered trademarks of CareFusion Corporation or one of its subsidiaries. All rights reserved.
All other trademarks are property of their respective owners.

© 2000-2012 CareFusion Corporation or one of its subsidiaries. All rights reserved.

This document contains proprietary information of CareFusion Corporation or one of its subsidiaries, and its receipt or possession does not convey any rights to reproduce its contents, or to manufacture or sell any product described. Reproduction, disclosure, or use other than for the intended purpose without specific written authorization of CareFusion Corporation or one of its subsidiaries is strictly forbidden.

CareFusion Switzerland 317 Sarl,
A-One Business Centre, Z.A Vers –La-Pièce n° 10, CH-1180, Rolle

EC REP CareFusion UK 305 Ltd., The Crescent, Jays Close, Basingstoke, Hampshire, RG22 4BS, UK

1000DF00327 Issue 3

carefusion.com

CareFusion